

Certificate of Compliance

Certificate: 1660426

Master Contract: 175074

Project: 1660426

Date Issued: April 29, 2005

Issued to: Vibro-Meter SA
Rte de Moncor 4
P.O. Box 1071
CH-1701 Fribourg
SWITZERLAND

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'

Issued by: O. Ewanchyna, P. Eng.



Authorized by: M.H.J. Hoendervangers



PRODUCTS

CLASS 8721 05 - LABORATORY ELECTRICAL EQUIPMENT

CLASS 8721 85 - ELECTRICAL EQUIPMENT FOR LABORATORY USE (Certified to U.S. Standards)

Machinery Protection and Condition Monitoring System, model VM600-CSA, 24Vdc – 400W, 48Vdc – 400W, 72Vdc – 400W, 110Vdc – 400W or 115 / 230Vac, 50/60 Hz – 400W

NOTE: The above model is Equipment Class I, Pollution Degree 2, Installation Category I

APPLICABLE REQUIREMENTS

- | | | |
|---|---|--|
| CAN/CSA-C22.2 No. 61010.1-04 | - | Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements |
| UL Std. No. 61010-1 (2 nd Edition) | - | Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |
| IEC 61010-1 (2 nd Edition) | - | Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements |

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards 166042 April 29, 2001 75071660426

Certificate: 1660426

Master Contract: 175074

Project: 1660426

Date: April 29, 2005

Conditions for use

- This unit has been evaluated for rack mounting. A suitable Electrical and Fire enclosure shall be provided in the end-product.

- Ambient temperature within range -25°C to 48°C , maximum allowable ambient temperature depends on configuration, needs to be evaluated according instruction manual (IM) clause 8.3.1 and appendix A.



Supplement to Certificate of Compliance

Certificate: 1660426

Master Contract: 175074

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
1660426	April 29, 2005	Original Certification of VM600-CSA